

Immucor, Inc.
 3130 Gateway Drive
 Norcross, GA 30091-5625
 Tel: (770) 441-2051 Toll-free: (800) 510-5110
 FAX: (770) 441-3807
 Web site <http://www.immucor.com>

Certificate of Analysis

This is to certify that the product listed hereunder has been manufactured and tested by Immucor, Inc., of Norcross, Georgia. The product listed has been tested in accordance with current requirements for the product in question, and has been found to meet all specifications.

Product: Weak D Cells

Catalog Nos. 0002995 / 1 x 5 mL

Lot No. 52160

Expiration date: 2018-03-02 **Date of manufacture:** 2017-12-25

Chemical Composition: Weak D human red blood cells. These red cells are suspended to 2-4% concentration in buffered preservative solution containing Neomycin Sulfate 0.1 mg/mL, Chloramphenicol 0.25 mg/mL, and Gentamycin Sulfate 0.05 mg/mL.

Viral Marker Testing:	Requirement	Non-reactive
	Source material nonreactive for antibodies to human immunodeficiency virus types 1 and 2 (anti-HIV) and human hepatitis C virus (Anti-HCV), and hepatitis B surface antigen (HBsAg); and nonreactive for HIV-1 RNA and HCV RNA by NAT using FDA-licensed tests and nonreactive for serologic test for syphilis (STS).	HBsAg <input checked="" type="checkbox"/> Anti-HCV <input checked="" type="checkbox"/> Anti-HIV <input checked="" type="checkbox"/> HIV-1 RNA <input checked="" type="checkbox"/> HCV RNA <input checked="" type="checkbox"/> STS <input checked="" type="checkbox"/>

Reactivity: These cells are expected to show little or no direct agglutination when tested with Anti-D reagents, yet show a definite positive reaction at the antiglobulin phase.

Identity testing: Product tested for identity as labeled.

Approved for Release: All specifications met; labeling checked and confirmed. **Released**

QC Reviewed by/Date: Frank Hugg 16 JAN 2018 Approved by/Date: B. B. [Signature] 18 JAN 2018