

**Immucor, Inc.**  
3130 Gateway Drive  
Norcross, GA 30091-5625  
Tel: (770) 441-2051 Toll-free: (800) 510-5110  
FAX: (770) 441-3807  
Web site <http://www.immucor.com>

## Certificate of Analysis

This is to certify that the product listed hereunder has been manufactured and tested by Immucor, Inc., of Norcross, Georgia. The product listed has been tested in accordance with current FDA requirements for the product in question, and has been found to meet all specifications.

**Product:** Capture LISS – Low Ionic Strength Solution For Capture Solid Phase Antibody Detection Tests

**Catalog No.** 0006420 / 10 x 11.5 mL

**Lot No.** 211011

**Expiration date:** 2018-09-26      **Date of manufacture:** 2017-10-09

**Chemical Composition:** Immucor LISS is a low ionic strength solution containing glycine and the dye bromcresol purple and the preservative sodium azide (0.1%).

**Viral Marker Testing:** Not applicable. Product contains no human source material.

Potency:	Requirement	Phenotype	Results
	No US standard for potency.	K+k+	8
		Fy(a+b+)	16

Specificity:	Requirement	Confirmed
	Product specificity is demonstrated by testing inert sera with unsensitized red cells and platelets by insert methods.	<input checked="" type="checkbox"/>
		<b>Released</b> <input checked="" type="checkbox"/>

**Approved for Release:** All specifications met; labeling checked and confirmed.

QC Reviewed By/Date: *Trish R 10 OCT 2017*      Approved By/Date: *D. Penfield 25 OCT 2017*