


# epoc Blood Analysis System Product Specifications

## Be where patients need you most.

Make patient-side testing the  
nexus of care hospital-wide.

[siemens-healthineers.ca/epocnxs](http://siemens-healthineers.ca/epocnxs)



### System Description

Patient-side blood gas analyzer

### System Menu

pH	Na <sup>+</sup>	Hct	BUN
pCO <sub>2</sub>	K <sup>+</sup>	Glu	Urea
pO <sub>2</sub>	Ca <sup>++</sup>	Lac	
TCO <sub>2</sub>	Cl <sup>-</sup>	Crea	

### Measured Parameters

Parameter	Unit of Measure	Measurement Range
pH	pH units	6.5–8.0
pCO <sub>2</sub>	mmHg kPa	5–250 0.7–33.3
pO <sub>2</sub>	mmHg kPa	5–750 0.7–100
TCO <sub>2</sub>	mmol/L mEq/L	5–50 5–50
Na <sup>+</sup>	mmol/L mEq/L	85–180
K <sup>+</sup>	mmol/L mEq/L	1.5–12.0
Ca <sup>++</sup>	mmol/L mg/dL mEq/L	0.25–4.00 1.0–16.0 0.5–8.0
Cl <sup>-</sup>	mmol/L mEq/L	65–140
Hct	% PCV L/L	10–75 0.10–0.75

### Measured Parameters continued

Parameter	Unit of Measure	Measurement Range
Glu	mmol/L mg/dL g/L	1.1–38.5 20–700 0.20–7.00
Lac	mmol/L mg/dL g/L	0.30–20.00 2.7–180.2 0.03–1.80
Crea	mg/dL μmol/L	0.30–15.00 27–1326
BUN	mg/dL	3–120
Urea	mmol/L mg/dL g/L	1.1–42.8 7–257 0.07–2.57

*Institutions should establish and set their own normal range values.*

### Calculated Parameters

Parameter	Unit of Measure	Measurement Range
cHgb	g/dL mmol/L g/L	3.3–25 2.0–15.5 33–250
cHCO <sub>3</sub> <sup>-</sup>	mmol/L mEq/L	1–85
cTCO <sub>2</sub>	mmol/L mEq/L	5–50
BE(ecf)	mmol/L mEq/L	-30–+30
BE(b)	mmol/L mEq/L	-30–+30

### Calculated Parameters continued

Parameter	Unit of Measure	Measurement Range
cSO <sub>2</sub>	%	0–100
GFRmdr*†	mL/min/ 1.73m <sup>2</sup>	2–60 or >60
GFRmdr-a*†	mL/min/ 1.73m <sup>2</sup>	2–60 or >60
GFRckd <sup>‡</sup>	mL/min/ 1.73m <sup>2</sup>	1–225
GFRckd-a <sup>‡</sup>	mL/min/ 1.73m <sup>2</sup>	1–225
GFRswz <sup>§</sup>	mL/min/ 1.73m <sup>2</sup>	1–275
AGap	mmol/L mEq/L	-14–+95
AGapK	mmol/L mEq/L	-10–+99
BUN/Crea	mg/mg	0.2–400.0
Urea/Crea	mmol/ mmol mg/mg	0.8–1615.4 0.4–856.8
A	mmHg kPa	5–800 0.67–106.64
A-a	mmHg kPa	1–800 0.13–106.64
a/A	% fraction	0–100 0–1

# epoc® Blood Analysis System Product Specifications (continued)

## Sample Types

Whole blood: arterial, venous, capillary

## Sample Size

At least 92 µL

## Time to Result

Less than 1 minute

## Calibration

Automatically performed prior to every test

## Quality Control

Eurotrol GAS-ISE Metabolites (with BUN)  
Eurotrol Calibration Verification Fluids (with BUN)  
Eurotrol Hct Control  
Eurotrol Hematocrit Verification Fluid

## Integrated Bar-code Scanner

For Patient ID, Operator ID, and ampuled QC;  
1D and 2D bar-code symbologies

## External Interfaces

HL7

## Power Requirements

100–240 VAC; 50–60 Hz

## Operation on Battery

At least 50 test cards

## Safety

IEC 61010-1, IEC 61010-2-81,  
IEC 61010-2-101

## EMC

IEC 61326-1, IEC 61326-2-6

## Data Security

Patient Data encryption  
Encrypted end-to-end data transmission  
No hard-coded password  
No user access to operating system

## Hardware

Operating System	Android 9.0
Processor	MT6762 8*Cortex-4*A53 2.0GHz/4*A53 1.5GHz
Memory	2 GB RAM; 16 GB ROM

## BLUETOOTH

BT4.2(BLE HS compliant) BT Smart Ready

## BLUETOOTH Security

128-bit encryption with PIN authentication

## WIFI

MediaTek 6631 chipset  
802.11 a/b/g/n/ac/r  
2.4GHz +5GHz\*\*

## WIFI Security

WEP, WPA/WPA2 Personal (Pre-Shared Key)  
EAP (802.1x): TLS, TTLS (MS-CHAP), TTLS (MS-CHAP v2), TTLS (CHAP),  
TTLS-MD5, TTLS-PAP, PEAP-TLS, PEAP (MS-CHAP v2), LEAP,  
FAST WAPI PSK, WAPI Certificate

## Data Capacity

Patient Test	499
QA Tests	2000
Operators	4000

## Test Card

Storage	15–30°C
Shelf Life	168 Days or Up to 6 months
Size	3.39 in (L) × 2.13 in (W) × 0.06 in (H) 86 mm (L) × 54 mm (W) × 1.4 mm (H)

## System Dimensions

	Host	Reader
Height (nominal)	160 mm [6.30 in]	215 mm [8.46 in]
Width (nominal)	78 mm [3.07 in]	85 mm [3.35 in]
Thickness (nominal)	16 mm [0.63 in] (22 mm [0.87 in] at scanner)	50 mm [2 in]
Weight (nominal)	250 g [8.82 oz] (with battery)	<500g [<1.1 lb]
Display	5.0" diagonal, HD 1280 × 720 pixels, touchscreen	—

## Environmental Requirements

Operating Temperature	-20° C to 55° C [-4° F to 130° F] (Host)	15°C–30°C (Reader and Test Card)
Humidity	Up to 85% relative humidity, non-condensing	
Barometric Pressure	400–825 mmHg (53.33–110 kPa)	

epoc, POCcelerator, and all associated marks are trademarks of Siemens Healthcare Diagnostics Inc., or its affiliates. All other trademarks and brands are the property of their respective owners.

Product availability may vary from country to country and is subject to varying regulatory requirements. Please contact your local representative for availability.

*\*Values >60 will be reported as  
>60 mL/min/1.73 m<sup>2</sup>  
†IDMS-traceable MDRD type  
‡CKD-EPI equation  
§Bedside Schwartz equation  
\*\*5GHz Wi-Fi Channel might not  
be available in some countries*

## Siemens Healthineers Headquarters

Siemens Healthcare GmbH  
Henkestr. 127  
91052 Erlangen, Germany  
Phone: +49 9131 84-0  
siemens-healthineers.com

## Legal Manufacturer

Epocal, Inc.  
2060 Walkey Road  
Ottawa, ON K1G 3P5  
Canada

## Distributed by

Siemens Healthcare Limited - Canada  
1577 North Service Road East  
Oakville, ON, L6H 0H6, Canada  
Phone: +1-905-465-8000  
siemens-healthineers.ca

## for the Product

epoc Blood Analysis System